

D-Dimer	D-Dimer
---------	---------


Sample No	22.35		
Sample Details	Plasma with a slightly elevated D-Dimer level		
Prior Use	None		
Unit	mg/L D-Dimer, mg/L FEU		
Expiry Date	19-December-2023		
Homogeneity	0.0 %	Homogeneity Parameter	D-Dimer
Number of Participants	731		
Number of Responders	697	Response Rate	95 %
Comments	The following participants reported deviating results which were excluded from the statistical evaluation: <b>198</b> : 0.017 <b>4545</b> : 0.077		

**Own Reagent**

	n	assigned value	Uncert.	CV (%)	Range	your result	z-score	your result	z-score	your result	z-score
IL HemosIL D-Dimer	9	0.62			0.38 - 0.70						
IL ACL 100 - 9000	1	0.50									
IL ACL Elite (Pro)	2	0.53			0.50 - 0.55						
IL ACL TOP Family	6	0.65			0.38 - 0.70						
IL HemosIL D-Dimer HS500	173	0.37	0.00	13.5	0.21 - 0.67	0.33	-0.72				
IL ACL TOP Family	172	0.37	0.00	13.5	0.21 - 0.67	0.33	-0.73				
Technoclone Ceveron	1	0.32									

**Other Reagents (mg/L D-Dimer)**

	n	assigned value	Uncert.	CV (%)	Range
BioMedica Diagnostics Dimertest	1	0.32			
IL HemosIL D-Dimer HS	27	0.12	0.01	22.0	0.08 - 0.23
Medirox D-Dimer	5	0.09			0.07 - 0.14
Nordic Red D-Dimer	1	0.07			
Other	2	0.21			0.16 - 0.27
Radiometer AQT90 Flex D-Dimer	12	0.20	0.01	11.0	0.17 - 0.28
Sysmex LIAS Auto D-Dimer Neo	12	0.77	0.02	8.5	0.65 - 0.85
Tulip Quantia D-Dimer	1	0.48			



**Other Reagents (mg/L FEU)**

	n	assigned value	Uncert.	CV (%)	Range
Beckman Coulter D-Dimer	6	0.18			0.10 - 0.25
BioMerieux Vidas D-Dimer	24	0.24	0.01	12.4	0.17 - 0.30
Boditech AFIAS / i-Chroma D-Dimer	3	0.13			0.13 - 0.16
Diagon Dia-D-Dimer	6	0.26			0.18 - 0.32
IL HemosIL D-Dimer 500	2	0.39			0.34 - 0.44
Mindray D-Dimer	17	0.50	0.02	12.7	0.29 - 0.74
Mitsubishi Pathfast D-Dimer	10	0.30	0.01	6.3	0.27 - 0.32
Quick D-dimer PLUS	1	0.20			
Roche Cardiac D-Dimer	12	0.10	0.00	0.0	0.10 - 0.13
Roche Tinaquant 2nd gen. (cal. citrate)	16	0.24	0.04	50.2	0.12 - 0.43
Roche Tinaquant 2nd gen. (cal. heparin)	33	0.17	0.01	21.7	0.08 - 0.40
Siemens Immulite D-Dimer	4	0.24			0.20 - 0.34
Siemens Innovance D-Dimer	384	0.77	0.01	12.9	0.42 - 1.75
Stago / Roche Asserachrom D-Dimer	1	0.18			
Stago / Roche Liatest D-Dimer	17	0.32	0.01	10.9	0.27 - 0.36
Stago Liatest D-Dimer Plus	102	0.32	0.01	17.6	0.15 - 0.63
Tosoh D-Dimer S	2	0.16			0.15 - 0.16
Tulip Quantia D-Dimer	2	0.57			0.52 - 0.61

